

Lab. ID #
117928**To:** Town of Sherborn
19 Washington St.
Sherborn, MA 01770

Received Date/Time:	7/7/25, 2:20 PM
Analyzed Date/Time:	7/7/25, 2:40 PM
Reported Date:	7/8/25
Collected By:	G & L Labs.
Collection Date:	7/7/25

ANALYTICAL RESULTS

Sample#	Sampling Location	Sampling Time	E. coli (CFU/100mL)
1	Farm Pond	10:50 AM	< 5
Maximum Contaminant Level			235
Method of Reference			EPA 1603.1

Laboratory QC Results

QC	Unit	E. coli
Duplicate (D1)	CFU/100mL	< 5
Duplicate (D2)	CFU/100mL	< 5
Log Dif. of Dup		0.00
MS Recovery	%	90%
MS Rec. Criteria	%	17%-117%
Dilution Blk.(Beginning)	CFU/100mL	Negative
Dilution Blk.(Ending)	CFU/100mL	Negative
Positive Control	CFU/100mL	Positive

Approved By:_____
Lab. Director_____
Date 7/8/25



CHAIN OF CUSTODY RECORD

LAB I.D. #:

117928

246 Arlington Street, Quincy, MA 02170
Tel: (617) 328-3663 Fax: (617) 472-0706

COMPANY: Town of Sherborn
ADDRESS: 19 Washington St.
Sherborn, MA 01770
PHONE #: 508-651-7851 FAX #: _____
P.O. #: _____
CLIENT CONTACT: Jeanne Guthrie
PROJECT ID/LOCATION: Farm Pond, Sherborn

SAMPLE TYPE	CONTAINER TYPE
1 WATER	P - PLASTIC
2 SOIL	G - GLASS
3 SLUDGE	V - VOA
4 OIL	
5 TISSUE	
6 DRINKING WATER	
OTHER	

PROJECT ID/LOCATION: Farm Pond, Sherborn

COMMENTS

ANALYSES

E. mail:

Sample Collected By:

RELINQUISHED BY:

DATE: 7 - 7 - 25

TIME: 2 - 30 - 30

RECEIVED BY:

DATE: 7-7-25

TIME: 2 - 20 min

1/

TIME: 2 - 20 - min

REINQUISHED BY:

DATE:

RECEIVED BY:

DATE: _____

RELINQUISHED BY:

DATE	-	-	-
------	---	---	---

RECEIVED FOR LIBRARY.

DATE: - -

METHOD OF SHIPMENT

☐ Yes ☐ No

If Yes, list known

SPECIAL INSTRUCTIONS

☒ **RUSH,** BUSINESS DAY TURNAROUND
☐ **ROUTINE**

Sample Disposal Information

Are there any other known or suspected contaminants in these samples other than those listed above ?

If Yes, list known

☐ No ☐ Yes

